

Features

- Easy operation by one button.
- End point, kinetic, fixed time, multistandard, bichromatic, etc.
- Filter wavelengths: 340/405/492/510/546/578/620nm, 2 more open filter positions, others on request.
- With 20 incubating positions.
- Large memory to store 200 test programs and 1000 test results.
- Excellent Q.C function, Q.C chart can be stored, displayed and printed.
- 3 levels temperature: 25°C, 30°C, 37°C can be selected in flow cell and incubator.
- Real time graph can be displayed and printed.

Technical Specifications

Model	SA-10	SA-20
Reading Cuvette	Direct reading cuvette	Both through cell and direct reading cuvette
Incubating Positions	20 incubating positions	
	Light source: 6V, 10 W longlife halogen lamp	
Photometric System	Filters: 340/405/492/510/546/578/620nm, 2 more open filter positions, ± 2 nm pass-band	
	Wavelength accuracy: ± 2 nm	
	Measuring range: 0-2.500 O.D.	
Measuring System	Photometric linearity: $\pm 2\%$ from 0 to 2.000 O.D.	
	Photometric accuracy: $\pm 2\%$ from 0 to 2.000 O.D.	
	Drift < 0.005 O.D.	
	Carry over $\leq 1\%$	
Incubator Temperature	25°C, 30°C, 37°C RT	
Control	Precision: $\pm 0.1^\circ\text{C}$	
Display	5 inch color touch screen	
Printer	Built-in thermal printer	
Connection Port	RS-232 & USB	
Power Supply	AC 110V 60Hz / 220V 50Hz	
Dimensions	34cmX38cmX18cm	
Weight	8.5kg	

CLINDIAG...

The Reference for Quality and Service

CLINDIAG
 CLINDIAG SYSTEMS B.V.B.A.


SA-10/20

Chemistry Analyzer



ISO9001:2008 & ISO13485:2003



Headquarters

Add: Steenberg 66, Pollare/Ninove(9401), Belgium
 Tel: 0032 54 250 936
 Fax: 0032 54 243 058
 Website: <http://www.clindia.be>
 E-mail: marketing@clindia.be

U.S. Office

Add: 1351 S. Leavitt Ave, # 104 Orange City, Florida 32763 USA
 Tel: 001 386 456 1235
 Fax: 001 386 456 1237
 E-mail: info@clindiausa.com

India Office

Add: B-1177-78, G.D. Colony Mayur Vihar Phase-III, Delhi-110096
 PH: 0091 11 22618779
 Fax: 0091 11 22618780
 E-mail: contact@clindiaindia.com